



National Laboratory for HIV Reference Services
National HIV and Retrovirology Laboratories
National Microbiology Laboratory
Public Health Agency of Canada

HIV Serology Quality Assessment Program Summary for Panel HIVSER 2021Apr19

2021Apr19 HIV Serology Panel		
Panel Sample	True Status	Labs Reporting Incorrect Status
A	HIV-1 Ab Positive	
B	HIV-2 Ab Positive	HV03
C	HIV-1/2 Ag/Ab Negative	
D	HIV-1/2 Ag/Ab Negative	
E	HIV-1 Ag Positive	HV03 HV59

Summary of findings observed for the 2021Apr19 panel:

- 1) Participant HV03 mixed up samples B and E during the testing stage.
- 2) Participant HV59 did not recommend further action for sample E.
- 3) Participant HV23 was not able to participate.



National Laboratory for HIV Reference Services
National HIV and Retrovirology Laboratories
National Microbiology Laboratory
Public Health Agency of Canada

HIV Serology Quality Assessment Program

Final Report for Panel HIVSER 2021Apr19

Issued 2021-July-19

Introduction

The NLHRS distributed the 2020Oct30 panel and the 2021Apr19 panel on October 14, 2020. This final report is specific to the 2021Apr19 panel only and is publicly available; however, the identity of participants has not been disclosed. The deadline for results submission was April 19, 2021. The preliminary report was issued on May 5, 2021.

Panel Samples, HIV Test Kits, and Data Entry

- *Panel Composition:*
 - The 2021Apr19 panel consisted of five samples: two HIV negative (C, D), one HIV-1 Ab positive (A), one HIV-2 Ab positive (B), and one HIV-1 Ag positive (E). Samples A and B were diluted 1 in 2 with defibrinated human plasma (Basematrix 53, Seracare Life Sciences). Testing and characterization of the panel by the NLHRS prior to shipment are presented in Appendix 2. Panels were sent to 41 participants including the NLHRS on October 14, 2020.
- *HIV Test Kits*
 - Nine different assays were used by the 40 participants (excluding the NLHRS) who returned results (Appendix 3).
- *Data entry*
 - Results entry for this panel utilized an in-house developed website.

Homogeneity and stability

- The homogeneity and stability of the 2021Apr19 HIV serology panel was assessed by comparing the participants' results with the panel characterization results obtained by the NLHRS prior to the panel send-out.
- There was no indication of heterogeneity or instability of the panel samples as the data submitted by the participants is consistent with the expected results from the NLHRS characterization of each panel member (Figures 1, 2, and Appendix 2).

Results

- *Evaluation Criteria:*

- Negative samples: HIV non-reactive/negative in the final HIV serology interpretation with assay results supporting the final serology interpretation.
- Positive samples: HIV reactive/positive in the final HIV serology interpretation with assay results supporting the final serology interpretation. Participants must provide a recommendation for further action for samples that they could not determine the true serology status for based on the assay used in their testing.

- *Qualitative Group Analysis (Figures 1 and 2)*

- *Sample A (HIV-1 Ab Positive)* – 41/41 participants provided either a correct serology status and/or recommendation.
- *Sample B (HIV-2 Ab Positive)* – 40/41 participants provided either a correct serology status and/or recommendation.
- *Sample C (HIV-1/2 Ag/Ab Negative)* – 41/41 participants provided either a correct serology status and/or recommendation.
- *Sample D (HIV-1/2 Ag/Ab Negative)* – 41/41 participants provided either a correct serology status and/or recommendation.
- *Sample E (HIV-1 Ag Positive)* – 39/41 participants provided either a correct serology status and/or recommendation.





Figure 1: The final HIV serology status of the positive samples in the 2021Apr19 HIV serology panel submitted by participants using an HIV screening assay.



*Waived flag as the participant was only assessed on the confirmatory assay portion of their testing algorithm that used both molecular and serological assays.

Figure 2: The final HIV serology status of the positive samples in the 2021Apr19 HIV serology panel submitted by participants (including NLHRS) using HIV screening and confirmatory assays.

Findings

The majority of participants correctly identified the serology status and/or provided an appropriate recommendation for the panel samples included in the 2021Apr19 test event.

One participant reported an HIV-2 Ab positive result for samples B (expected result: HIV-2 Ab positive) and E (expected result: HIV-1 Ag positive) as determined by the Geenius assay. The samples were returned to the NLHRS for retesting on both the Geenius and Bio-Rad Genscreen HIV-1 Ag assays. Both samples were found to be HIV-2 Ab positive. Additionally, sample B was found to be HIV-1 Ag reactive and sample E was found to be HIV-1 Ag non-reactive. These aberrant results were potentially caused by a sample mix up or contamination event that likely explains the deviation from the expected results (Appendix 2).

One participant did not provide an appropriate recommendation for further action for sample E.

In closing, we value each laboratory's participation in these QA test events and your suggestions for improvement. The NLHRS is committed to improve all aspects of the HIV serology proficiency testing program in order to provide quality proficiency testing to our participants.

If you have any comments, suggestion or concerns, please contact us at:

phac.nlhrs.qap-peq.lnsrv.aspc@canada.ca

Thank you for your participation in the NLHRS HIV Serology QA Program




John Ho
Quality Assurance Program Coordinator
National Laboratory for HIV Reference Services
Public Health Agency of Canada
Tel: (204) 789-6518




Dr. John Kim
Laboratory Chief
National Laboratory for HIV Reference Services
Public Health Agency of Canada
Tel: (204) 789-6527

Appendix 1 - CLSI M53 Algorithm I


(i) HIV-1/HIV-2 Ag/Ab Immunoassay



(ii) HIV-1/HIV-2 Ab Differentiation Immunoassay



(iii) Nucleic Acid Testing




Appendix 1: Adaptation of the Clinical and Laboratory Standards Institute (CLSI) M53-*Criteria for Laboratory Testing and Diagnosis of Human Immunodeficiency Virus Infection: Approved Guideline* Algorithm I.

Appendix 2: Summary of NLHRS characterization of the 2021Apr19 HIV serology panel samples

Sample		C/D (Duplicate)	A	B	E
Final HIV Status		HIV-1/2 Ag/Ab Negative	HIV-1 Ab Positive	HIV-2 Ab Positive	HIV-1 Ag Positive
bioLytical INSTI HIV-1/2 Rapid Test	Result	Non-Reactive	Reactive	Reactive	Non-Reactive
Bio-Rad GS HIV p24	Result	Non-Reactive	Non-Reactive	Non-Reactive	Reactive
Bio-Rad GS HIV p24 Confirmatory	Result	Not Tested	Not Tested	Not Tested	99.6% Neutralization
Fujirebio INNO-LIA HIV-I/II Score	Result	Negative	HIV-1	HIV-2	Negative
	sgp120	-	++	-	-
	gp41	-	++	-	-
	p31	-	++	++	-
	p24	-	+++	+/-	-
	p17	-	+++	-	-
	sgp105	-	-	+	-
Bio-Rad Geenius HIV-1/HIV-2 Supplemental Assay	Result	Negative	HIV-1	HIV-2	Negative
	gp36	-	-	+	-
	gp140	-	+	+	-
	p31	-	+	+	-
	gp160	-	+	-	-
	p24	-	+	-	-
	gp41	-	+	-	-
CTRL	+	+	+	+	

Appendix 3: Summary of assays used by the participants in the 2021Apr19 HIV serology test event



Appendix 4: Summary of bands detected for samples A, B, and E by the Bio-Rad Geenius HIV-1/2 confirmatory assay in the 2021Apr HIV serology test event

Bio-Rad Geenius	Frequency of Bands Detected						
	gp36	gp140	p31	gp160	p24	gp41	CTRL
2021Apr19A	-	10	12	11	11	11	11
2021Apr19B	11	11	11	-	-	-	11
2021Apr19E	1	1	1	-	-	-	11

Appendix 5: Troubleshooting

Troubleshooting; common causes of outlying and/or aberrant results in serology and molecular Laboratories.

Type of Error	Possible Cause(s)	Pre-Analytical	Analytical	Post- Analytical
Sample mix-up	Can occur during specimen reception or testing. May result in outlying/aberrant results for one or all samples mixed-up.	✓	✓	
Transcription	• Incorrect test ordering by physician	✓		
	• Incorrect shipment address	✓		
	• Selecting the wrong assay for data entry	✓		
	• Interchanging results for two or more specimens			✓
	• Entering incorrect results			✓
	• Entering values in the incorrect field (e.g., OD as S/Co)			✓
	• Entering values in the incorrect unit (e.g., IU/mL instead of log ₁₀ copies/mL)			✓
	• Using a comma instead of a dot to denote a decimal point			✓
	• Selecting the incorrect assay interpretation or analyte			✓
	• Failure to recommend follow-up testing where necessary			✓
It is recommended all results that are manually transcribed or entered electronically be checked by a second individual to avoid transcription errors.				
Outlying and/or Aberrant Results (<u>random error</u>)	<u>Sporadic test results identified as outlying and/or aberrant can be classified as random events. Possible causes of random error include:</u>			
	• Incorrect sample storage/shipping conditions	✓	✓	
	• Incorrect test method	✓	✓	
	• Insufficient mixing of sample, especially following freezing		✓	
	• Poor pipetting		✓	
	• Ineffective or inconsistent washing		✓	
	• Transcription errors	✓		✓
	• Cross-contamination or carryover	✓	✓	
• Presence of inhibitors to PCR		✓		
Outlying and/or Aberrant Results (<u>systematic error</u>)	<u>A series of test results identified as outlying and/or aberrant may be due to a systematic problem. Systematic problems may be due to:</u>			
	• Reagents contaminated, expired, or subject to batch variation		✓	
	• Instrument error or malfunction		✓	
	• Insufficient washing		✓	
	• Incorrect wavelength used to read the assay result		✓	
	• Cycling times too long/short or temperature too high/low		✓	
	• Incubation time too long/short or temperature too high/low		✓	
	• Insufficient mixing/centrifuging before testing		✓	
	• Incorrect storage of test kits and/or reagents	✓		
	• Contamination of master-mix, extraction areas or equipment		✓	
	• Ineffective extraction process		✓	
	• Degradation of master-mix components		✓	
• Suboptimal primer design (in-house assays)		✓		

This table was modified from a report produced by the National Reference Laboratory (NRL), Melbourne, Australia.